

SYNTHESIS DOCUMENT- EURO 1000
PRODUCTS CONFORMITY ASSESSMENT

1 SCOPE OF THIS DOCUMENT

The present document constitutes the synthesis of the APAVE report n° n° 6422931 which gives the details of inspections and tests achieved on the product (this document replaces the previous document of 20/02/2017 - validity period revised to 5 years).

2 CONCERNED PRODUCT SPECIFIED FEATURE

- | | | | |
|-----------------------|--|---------------------------|--------------|
| • Denomination : | Medical beds for adults | • Application environment | 2, 3, 4 et 5 |
| • Designation : | Product group Euro 1000 | • Tension : | 230 V |
| • Make : | HMS - VILGO | • power/current | 185 W |
| • Model : | Euro 1000 ; 1002 ; 1400 ; 1402 ; 1500 ; 1600 ; 1800 | • frequency : | 50 Hz |
| • Weight max patient: | 135kg | • Protective Indices | IP66 |

3 SCOPE OF THE MISSION (to see details in report)

Conformity Assessment of the product and its technical file to the following documents :
⇒ Standard NF EN 60601-2-52 of June 2010, , used with standard NF EN 60601-1 edition 3,
⇒ Standard NF EN 60601-1-2 edition 3,
⇒ Standard NF EN 60601-1-11 edition 1.

CONCLUSION OF THE REPORT

The conclusion of the report shows that the examined product is in compliance with the prescriptions of the reference documents.

Tassin the 14/02/2018

Limit of validity of this document : 02/2022

Charge of Development Missions
Testing Product Laboratory

Document original immatériel



Xavier SCHMITT

This document is not a report established under the title of a Notified body, it is a synthesis of the mission achieved by APAVE and only the complete French report maybe used as reference. It doesn't substitute itself to the DECLARATION OF CONFORMITY issued by the manufacturer, according the market regulation of the EUROPEAN UNION and cannot be use as a guarantee of the production conformity to the type.
Its reproduction is only allowed as a complete photographic facsimile